



Certificate of Sterility

Beta Treatment

LLG-Cryotubes

TREATMENT PARAMETERS

METHOD

Irradiation with accelerated electrons

ENERGY (MeV)

10

ROUTINE DOSIMETERS

DOSIMETER TYPE

FWT

BATCH

1131

DREF MIN/DREF MAX (kGy)

20,1 - 23,1

ROUTINE DOSIMETERS

21,1 - 20,6

CUSTOMER SPECIFICATIONS DOSE (kGy)

18,0 - 22,1

CALCULATED DOSE (kGy)

18,4 - 20,1

Lab Logistics Group GmbH
Product Management
Am Hambuch 1
D-53340 Meckenheim / Germany

Meckenheim, 2016-07-28

This document was sent online and is valid without signature

