

User Manual
English

Portavo® 904 OXY



Warranty

Defects occurring within 3 years from delivery date shall be remedied free of charge at our plant (carriage and insurance paid by sender).

Sensors and accessories: 1 year

Subject to change

Return of products under warranty

Please contact our Service Team before returning a defective device.

Ship the cleaned device to the address you have been given.

If the device has been in contact with process fluids, it must be decontaminated/ disinfected before shipment. In that case, please attach a corresponding certificate, for the health and safety of our service personnel.



Disposal

Please observe the applicable local or national regulations concerning the disposal of "waste electrical and electronic equipment".

Registered trademarks

The following names are registered trademarks. For practical reasons they are shown without trademark symbol in this manual.

- Calimatic®
- Memosens®
- Paraly®
- Portavo®
- Sensocheck®
- Sensoface®

Package Contents	6
Documentation	7
Overview of the Portavo 904 OXY	8
Value-Added Features	9
Protective Cover	10
Hook	10
Display	11
Keypad	12
Start-Up	13
Inserting the Batteries	13
Connecting a Sensor	14
Switching On the Meter	15
Icons	15
Configuring	16
Calibrating	17
Measuring	21
Adjusting the Temperature	21
Data Logger	22
Operating Modes of the Data Logger (Logger Type)	23
Data Logger Menu	25
Configuring the Data Logger	26
Starting the Data Logger using CONT	27
Starting the Data Logger using START	27
Displaying the Logger Data	28
Stopping the Data Logger	29
Clearing the Data Logger	29
Clock	30
Paraly SW 112 Software	31
Error and Status Messages	32
“Sensoface” Messages	33
Error Messages	34

Product Line 35
Sensors 35
Accessories 35
Specifications..... 36
Index 38

Check the shipment for transport damage and completeness.

The package of the Portavo 904 OXY includes:

- The Portavo 904 X OXY incl. 4 AA batteries and premounted quiver
- Carrying strap
- Quickstart instructions in various languages
- Specific test report
- Safety instructions
- Data carrier with detailed user manuals and Paraly SW 112 software
- USB cable, 1.5 m

Specific Test Report

CD-ROM

Complete documentation:

- User manuals in different languages
- Safety instructions
- Quickstart guides



Safety information

In official EU languages and others.

- EU Declarations of Conformity



Quickstart Guides

Installation and first steps:

- Operation
- Menu structure
- Calibration
- Error messages and recommended actions



Various languages on CD-ROM and on our website:
www.knick.de



The Portavo 904 OXY is a portable oxygen meter. A plain-text line on the high-contrast LCD screen makes operation virtually self-explanatory. The meter stands out by the following features:

- Use of digital Memosens sensors
- A detachable quiver protects the sensor and prevents it from drying out. Furthermore, it can be used for calibration.
- The rugged housing is made of a high-performance polymer. It provides high impact resistance and dimensional stability even when exposed to extreme moisture.

- Scratch-proof clear glass display, perfectly readable even after years
- Very long operating time with one set of batteries (4 x AA) or use of a Li-ion battery for reliable operation even at high or very low operating temperatures
- Data logger with 5000 values
- Micro USB port for communication with Paraly SW 112 software for data evaluation of digital sensors (Memosens)
- Sensoface icons provide single-glance information on the sensor condition (page 33)
- Real-time clock and indication of battery charging level

Value-Added Features

Memosens

The Portavo 904 can communicate with Memosens sensors. These digital sensors are automatically identified and the meter switches to the appropriate measurement method. When a Memosens sensor is connected to the meter, it is indicated by the logo shown on the right. Furthermore, Memosens allows the storage of calibration data, operating times and other data which will be available and can still be used when the sensor is connected to another Memosens-capable device.



Sensoface


Sensoface provides quick information on the sensor condition. The three "smiley" faces as shown on the right represent the sensor condition during measurement and after a calibration. When the condition deteriorates, an "INFO ..." message gives a hint to the cause.





Protective Cover

The front of the meter is protected by a cover, which can be completely flipped over and secured to the back for operation. A label on the inner side of the cover explains the control functions and device messages.



Hook

A fold-out hook on the back allows suspending the meter. This leaves your hands free for the actual measurement. The **rating plate** is located beneath the hook.




Protective Cover and Hook Combined

Cover and hook can be joined together to form a benchtop stand allowing comfortable and fatigue-free working at a lab bench or desk.

Display

The meter has a three-line display for representing alphanumeric information such as measurement and calibration data, temperatures and date/time. Additional information is provided by means of icons (Sensoface, battery icon, etc.).


Some typical displays are shown here.




Measuring
(display of measured value and temperature)



Calibration – step 1
(calibration method: in air)



Calibration – step 2
(adjusting the relative humidity)



Logger data
(display of measured value, memory location, temperature, date and time)



Clock
(display of hours and minutes, seconds and date)



Keypad

The keys of the membrane keypad have a noticeable pressure point.

They have the following functions:

on/off	Switches the meter on and displays the device and calibration data (see Start-up)
meas	Switches the meter on / Activates measuring mode / Stopping the data logger
cal	Starts calibration
set	Activates configuration / Confirms entries
clock	Displays time and date, allows setting the clock using set
RCL	View stored values
STO	Holds and saves a measured value, allows setting and starting the logger by pressing set (page 22)
▲▼	When this icon is displayed, you can use the arrow keys for navigation.

Check the shipment for transport damage and completeness (see Package Contents).



Caution!

Do not operate the device when one of the following conditions applies:

- the device shows visible damage
- the device fails to perform the intended function
- prolonged storage at temperatures above 70 °C
- severe transport stresses

In this case, a professional routine test must be performed.

This test should be carried out by the manufacturer.

Inserting the Batteries



With four AA batteries, the Portavo has an operating time of approx. 500 h. Open the battery compartment on the rear of the device. Be sure to observe the correct polarity when inserting the batteries (see markings in the battery chamber). Close the battery compartment cover and screw it handtight.

A special lithium-ion battery suited to the battery compartment is available for the Portavo 904. The battery is recharged through the USB port.

A battery icon in the display indicates the battery power level:

	Icon fully filled	Batteries at full capacity
	Icon partially filled	Battery capacity is sufficient
	Icon empty	Battery capacity not sufficient; calibration is possible
	Icon blinks	Max. 10 operating hours remaining, measurement is still possible Caution! It is absolutely necessary to replace the batteries.

Connecting a Sensor


The Portavo 904 OXY provides several connections so that many types of sensors can be used for measurement (see illustration below). Note that only **one** sensor may be connected to the meter at a time.

The meter recognizes the connected Memosens sensor and displays the Memosens logo.

Separate temperature probe

After power-on, a separate temperature probe is automatically recognized.

When you want to replace the temperature probe, you must switch off the meter and then switch it on again.



Connections

- a - Micro USB port
- b - M8, 4 pins, for Memosens lab cable
- c - Temperature probe GND
- d - Temperature probe
- e - M12, 8 pins, for Memosens sensors

Memosens sensors have a **cable coupling**, which allows convenient replacement of sensors while the cable remains connected to the meter. The connecting cable is connected to socket **b** (M8, 4 pins) or **e** (M12, 8 pins).





Switching On the Meter

When you have connected the sensor, you can switch the meter on by pressing the **on/off** or **meas** key.




When the meter is switched on with the **on/off** key, first a self test is performed and then the calibration data and settings are displayed before the meter switches to measuring mode.

When the meter is switched on with the **meas** key, it immediately switches to measuring mode.

Depending on the connected sensor and the specific measuring task, several steps for configuration and calibration must be performed as described on the following pages.

Icons

Important information about the state of the device:





Oxy Configuration

Prior to measurement, a configuration should be performed to match the connected sensor and the desired measurement performance. Furthermore, you can select the suitable calibration method. The following table gives you an overview. Factory settings are shown in **bold print**.

Measurement

↓
set

"Setup" display

▲	Display 1
	Display 2
	Altitude
	Salt Correct
	CAL
▲	CAL Timer
▼	Auto OFF
	Temp Unit
	Time Format
	Date Format
	Default

set
↔

Select using arrow keys, confirm by pressing **set**.

Saturation in % air Concentration in mg/l
OFF date + time date time
0 ... 4000 m
0.0 ... 45.0 g/kg
AIR CAL ZERO CAL DATA INPUT FREE CAL
OFF 1 ... 99 days
OFF 0.1h 1h 6h 12h
°C °F
24h 12h
dd.mm.yy mm.dd.yy
NO YES (reset to factory settings)
Note: All data logger entries will be deleted.


- ▲ This icon prompts you to select a menu item using the arrow keys –
- ▼ the selection is confirmed by pressing **set**.



AIR CAL Calibration

(Calibrating the slope in air)

The calibration method is selected in the configuration menu.




Note: To abort calibration, you can press **meas** at any time.



Zero CAL Calibration

(Zero calibration with oxygen-free medium)

The calibration method is selected in the configuration menu.



Note: To abort calibration, you can press **meas** at any time.



DATA INPUT Calibration

(Calibration by entering known sensor values)

The calibration method is selected in the configuration menu.

Measurement

↓ **cal**

CAL
DATA INPUT

↓

xx blinks
nA
ZERO POINT

Use ▲▼ to adjust the known value for the sensor zero point.

↓ **cal**

xxx blinks
nA
SLOPE

Use ▲▼ to adjust the known value for the sensor slope.

↓ **cal**

Calibration will be performed. Automatic return to measuring mode.

Note: To abort calibration, you can press **meas** at any time.



FREE CAL Calibration

(Free selection of calibration method)

FREE CAL calibration is selected in the configuration menu.

Measurement



CAL
AIR CAL blinks



Use ▲▼ to select the desired calibration method (AIR CAL, ZERO CAL, DATA INPUT).

Perform the selected calibration (see AIR CAL, ZERO CAL or DATA INPUT calibration).

Once you have completed all preparations, you can start with the actual measurement.

Keys for measurement

- 1) Connect the desired sensor to the meter. Some sensors require a special preparation. Please proceed according to the operating instructions for the sensor.
- 2) Switch the meter on using the **on/off** or **meas** key.
- 3) Depending on the measurement method and the sensor used, immerse the sensing part of the sensor in the medium to be measured.
- 4) Watch the display and wait for the reading to stabilize.
- 5) By pressing the **STO** key, you can hold and save a measured value (see data logger, page 22).



Measurement can also be controlled via the Paraly SW 112 software.

Adjusting the Temperature

When you connect a sensor without temperature detector, you can manually adjust the temperature for measurement or calibration:

- 1) Press **meas** to access measuring mode.
The adjusted temperature will be displayed.
- 2) Set the desired temperature value using the ▼ or ▲ arrow.
Holding the key depressed changes the temperature value at high speed.

The Data Logger

The meter provides a data logger. **Prior to use**, it must be configured and then activated. You can choose from the following logger types:


- DIFF (signal-controlled logging of measured variable and temperature)
- INT (time-controlled logging at a fixed interval)
- DIFF+INT (combined time- and signal-controlled logging)
- SHOT (manual logging by pressing the **STO** key)

The data logger records up to 5000 entries and saves them in a circular buffer. Already existing entries will be overwritten.

The following data are recorded: primary value, temperature, time stamp and device status.

The Paraly SW 112 software allows convenient management of the data logger. It is always the currently selected process variable which is recorded. The “STO” icon and the memory address is displayed briefly to indicate that an entry is being saved.

Display: Icons related to the data logger



Operating Modes of the Data Logger (Logger Type)

Manual logging when logger is activated (SHOT)

In this mode, a measured value is recorded when the **STO** key is pressed.

Measurement
Logger **activated**



STO

The measured value is saved to the address of the last recorded value + 1

Manual logging when logger is deactivated

Measurement
Logger **deactivated**



STO

Measured value is maintained
Proposed address blinks
(address of the last recorded
value + 1)

If desired: Select start address
using ▲▼.




STO

Measured value is saved to the desired address (e.g. for overwriting an incorrect measurement).

Interval (INT)


In this mode, the measured values are cyclically recorded.

Example: Logger type "INT"




Difference (DIFF)

When the delta range (process variable and/or temperature) related to the last entry is exceeded, a new entry is created and the delta range is displaced upwards or downwards by the delta value. The first entry is automatically created when the data logger is started.



Difference + Interval combined (DIFF+INT)


When the delta range related to the last DIFF entry is exceeded, a new entry is created (example: entry **A**) and the delta range is displaced upwards or downwards by the delta value. As long as the measured value remains within the delta range, logging is performed at the preset interval. The first DIFF entry is automatically created when the data logger is started.



Data Logger Menu

Logger display

Select using arrow keys, confirm by pressing **set**.



Overview of data logger menu (default in bold print)			
Logger type	DIFF	Delta % air	OFF 0.1 ... 100 % air 1.0 % air OFF 0.01 ... 20 mg/l 1.00 mg/l
		Delta °C / °F	OFF 0.1 ... 50.0 °C 1.0 °C OFF 0.1 ... 90 °F 1.0 °F
	INT	Interval	h:mm:ss 0:00:01 ... 9:59:59 0:02:00
	DIFF+INT	DIFF	See logger type DIFF
		INT	See logger type INT
	SHOT	Currently selected process variable is recorded	

Configuring the Data Logger

Prerequisite: The data logger is stopped (press **meas**).

Measurement

↓ **STO**

Measured value is maintained

↓ **set**

Logger: CONT blinks

↓ ▼

Logger: START blinks

↓ ▼

Logger: DEL blinks

↓ ▼

Logger: SET blinks

↓ **set**

Logger: Current logger type
blinks


↓ **set**

Select desired logger type using ▲▼:
DIFF, INT, DIFF+INT or SHOT.

Select the appropriate parameters using ▲▼ and confirm each selection by pressing **set**. When configuration is finished, CONT blinks. You can start the data logger by selecting START or CONT (see page 27).

Starting the Data Logger using CONT

Prerequisite: Data logger is configured. Every time the meter has been switched off, the data logger must be restarted (exception: SHOT).



If desired: Select start address using ▲▼.


The measured value is saved to the selected start address (exception: SHOT).

“... FREE MEMORY” is displayed.

“LOGGER” and “active logger type” icons are displayed.

Starting the Data Logger using START

Prerequisite: Data logger is configured. All existing entries are deleted. The start address for saving the values is 0001. Every time the meter has been switched off, the data logger must be restarted (exception: SHOT).



All entries will be deleted. “5000 FREE MEMORY” is displayed.

“LOGGER” and “active logger type” icons are displayed.

Displaying the Logger Data

Pressing the **RCL** key displays all stored values. The Paraly SW 112 software allows convenient management of the data logger.

Measurement


RCL

The "RCL" icon and the last recorded value is displayed.


Use ▲▼ to select the desired address. Empty memory locations will also be displayed.

RCL or meas

Return to measurement



Example:
Measured value stored at location
0026



Example:
Empty memory location 0004

Stopping the Data Logger

You can stop the data logger at any time by pressing the **meas** key.

Measurement, logger **activated**



meas

Data logger is stopped. "LOGGER" and "active logger type" icons are no longer displayed. It is still possible to hold a measured value by pressing **STO** and send it to any desired address.

Clearing the Data Logger

Selecting "DEL" deletes all data records.

Measurement



STO

Measured value is maintained



set

Logger: CONT blinks



Logger: START blinks



Logger: DEL blinks
PRESS SET



set

All stored data are deleted.
"0000 DELETED" is displayed.



Press the **clock** key to access the clock mode. Date and time will be displayed in the format as set in the configuration menu.
To set the clock, proceed as follows:

Display of
time+date

↓ **set**

Hour display blinks
SET HOUR



Set value.

↓ **set**

Minute display blinks
SET MINUTE



Set value.

↓ **set**

Second display blinks and
shows 00

set

Clock is started, the seconds count up.

↓ **set**

Year display blinks
SET YEAR



Set value.

↓ **set**

Month display blinks
SET MONTH



Set value.

↓ **set**

Day display blinks
SET DAY



Set value.

↓ **set**

Display of
corrected time+date


The Paraly SW 112 software supplements the Portavo series. It allows convenient management of the data that have been acquired by the meters as well as simple and clear configuration of the meters. Paraly SW 112 starts automatically when the Portavo USB port is connected to the computer.

The Paraly SW 112 software stands out by the following features:


- Intuitive Windows user interface
- Easy configuration and management of several meters
- Display of device and sensor information
- Convenient management and evaluation of the data logger
- Export function for Microsoft Excel
- Print function
- Updating the device software

Note: A detailed user manual for the Paraly SW 112 software can be found on the included data carrier.

Error messages are indicated as “ERROR ...” on the display. Information on the sensor condition is indicated by the “Sensoface” icon (friendly, neutral, sad) possibly accompanied by an info message (“INFO ...”).



Example of an error message:
ERROR 1 (value out of range)



Example of a “Sensoface” message:
INFO 1 (cal timer expired)

Sensoface (the “smiley” icon) provides information on the sensor condition (main-tenance request). Measurement can still be performed. After a calibration, the corre-sponding Sensoface icon (friendly, neutral, sad) is shown together with the calibration data. Otherwise, Sensoface is only visible in measuring mode.

The most important error messages and “Sensoface” info messages are shown on the inside of the protective cover. A com-plete list of messages and their meanings is provided in the following tables.



“Sensoface” Messages

The “Sensoface” icon provides information on the sensor condition:

Sensoface	Meaning
-----------	---------



Sensor is okay



Calibrate the sensor soon



Calibrate or replace the sensor

The “neutral” and “sad” Sensoface icons are accompanied by an “INFO ...” message to give a hint to the cause of deterioration.

Sensoface	Message	Cause
 	INFO 1	Calibration timer
	INFO 5	Zero / Slope
	INFO 6	Response time
	INFO 8	Leakage current

Error Messages

The following error messages can be shown in the display.

Message	Cause	Remedy
 blinks	Battery empty	Replace the batteries.
ERROR 1	Value out of range	Check whether the measurement conditions correspond to the adjusted measuring range.
ERROR 3	Temperature value out of range	
ERROR 4	Zero point too high/low	Thoroughly rinse the sensor and recalibrate. If this does not help, replace the sensor.
ERROR 5	Slope too high/low	
ERROR 11	Measured value unstable Stability criterion not met	Leave the sensor in the liquid until the temperature is stable. If this does not help, replace the sensor.
ERROR 14	Time and date invalid	Set time and date
ERROR 18	Configuration invalid	Restart, reset to factory settings (Setup: DEFAULT YES), configure and calibrate. If this does not help, send in the device for repair.
ERROR 19	Factory settings error	Device defective, send it in.
ERROR 21	Sensor error (Memosens)	Connect operational Memosens sensor.
ERROR 22	Sensor conflict	Connect only one sensor.

Sensors

Digital oxygen sensors

Amperometric oxygen sensor (Memosens)

Order No.

SE 315 MS

Temperature detectors

Pt 1000 temperature detector

ZU 6959

Pt 1000 temperature detector with tilted tip

ZU 0156

Accessories

Item

Order No.

Robust field case (for meter, sensor,
various small parts and user manual)

ZU 0934

Replacement quiver (5 units)

ZU 0929

Memosens lab cable, M8, 4 pins

CA/MS-001XFA-L

Flexible connecting cable for Memosens sensors (M12, 8-pin)

CA/MS-001XDA-L

Li-ion battery

ZU 0925

O₂ membrane kit

ZU 0564

(4x membrane module, O-ring set, 25 ml electrolyte)

O₂ electrolyte

ZU 0565

Please visit our website for more information on our product range: www.knick.de.

Memosens input, oxygen	M8 socket, 4 pins, for Memosens lab cable or M12 socket, 8 pins, for Memosens sensors
Display ranges ¹⁾	Saturation 0.000 ... 200.0 % Concentration 000 µg/l ... 20.00 mg/l
Temperature meas. range ¹⁾	-20 ... +150 °C
Sensor standardization	Automatic calibration in air (100 % RH) Zero calibration
Connections	1 x M8 socket, 4 pins, for Memosens lab cable 1 x M12 socket for Memosens sensors 2 x 4-mm socket for separate temperature detector 1 x micro USB-B for data transmission to PC
Display	LCD STN 7-segment display with 3 lines and icons
Sensoface	Status indication (friendly, neutral, sad)
Status indicators	For battery power level, logger
Notices	Hourglass
Keypad	[on/off], [cal], [meas], [set], [▲], [▼], [STO], [RCL], [clock]
Data logger	With up to 5000 memory locations
Recording	Manual, interval- or event-controlled
Communication	USB 2.0
Profile	HID, driverless installation
Usage	Data exchange and configuration via Paraly SW 112 software
Diagnostics functions	
Sensor data (Memosens only)	Manufacturer, sensor type, serial number, operating time
Calibration data	Calibration date, zero, slope
Device self-test	Automatic memory test (FLASH, EEPROM, RAM)
Device data	Device type, software version, hardware version
Data retention	Parameters, calibration data > 10 years
EMC	EN 61326-1 (General Requirements)
Emitted interference	Class B (residential area)
Immunity to interference	Industry EN 61326-2-3 (Particular Requirements for Transmitters)

¹⁾ Ranges depending on Memosens sensor

RoHS conformity	According to directive 2011/65/EU
Power supply	
Portavo 904	4 x AA cells or 4 x NiMH rechargeable batteries 1 x Li-ion battery, USB chargeable
Operating time	Approx. 500 h (alkaline)
Nominal operating conditions	
Ambient temperature	-10 °C ...+55 °C
Transport/ Storage temperature	-25 ... +70 °C
Relative humidity	0 ... 95 %, short-term condensing allowed
Housing	
Material	PA12 GF30 (silver gray RAL 7001) + TPE (black)
Protection	IP 66/67 with pressure compensation
Dimensions	Approx. (132 x 156 x 30) mm
Weight	Approx. 500 g

0000 DELETED ("data deleted" display) 29

A

AA batteries 13

Accessories 35

AIR CAL (calibration) 17

Altitude (configuration) 16

Amperometric oxygen sensor (Memosens) 35

Arrow keys 12

B

Battery capacity 13

Battery charge indicator 13

Battery compartment 13

Battery icon 13

Battery replacement 13

Benchtop stand 10

C

Calibration (AIR CAL) 17

Calibration (DATA INPUT) 19

Calibration (FREE CAL) 20

Calibration (ZERO CAL) 18

cal key 12

CD-ROM 7

Charge level of batteries 13

Clearing the datalogger 29

Clock 30

clock key 12

Configuring the data logger 26

Connecting a sensor 14

Connecting cable for Memosens 14

Connecting cable for Memosens, order no. 35

Connections 14

Connection, USB 13

Continuous recording of measured values 23

CONT, starting the data logger 27

Cyclic recording of measured values 23

D

- DATA INPUT (calibration) 19
- Data logger 22
- Data logger configuration 26
- Data logger icons 22
- Data logger menu 25
- Data memory 22
- Data of the meter 36
- Date 30
- Default (configuration) 16
- Deleting data logger entries 29
- Delta range (data logger) 24
- Device configuration 16
- Device messages 32
- Device properties 8
- Difference (data logger mode) 24
- Difference+Interval (data logger mode) 24
- Display 11
- Display icons 15
- Displaying the data logger 22
- Displaying the time and date 30
- Disposal 3
- Documentation 7

E

- Electrolyte (accessory) 35
- ERROR (error codes) 34
- Error messages 32
- Error messages, overview 34
- EU Declarations of Conformity 7

F

- Features 8
- Field case (accessory) 35
- Flexible connecting cable for Memosens sensors (M12, 8-pin) 35
- FREE CAL, calibration method selected in measuring mode 20

H

Holding the measured value 23

Hook 10

Hours, display 30

I

Icons 15

Icons for data logger 22

INFO messages 33

Inserting the batteries 13

Interrupting the data logger 29

Interval (data logger mode) 23

Introduction 8

K

Keypad 12

L

Li-ion battery (accessory) 35

Lithium-ion battery 13

Logger 22

Logger display 25

Logger type (data logger modes) 23

M

meas key 12

Measured-value recording 23

Measuring 21

Membrane kit (accessory) 35

Memory for measured values 22

Memosens 9

Memosens connecting cable 14

Memosens lab cable (accessory) 35

Memosens sensors 14

Menu of data logger 25

Menu structure of configuration 16

Menu structure of data logger 25

Messages 32

Messages, "Sensoface" 33

Micro USB port 8
Minutes, display 30

N

Nitrogen 5.0 18

O

O2 electrolyte (accessory) 35
O2 membrane kit (accessory) 35
on/off key 12
Operating modes of the data logger 23
Order numbers (accessories) 35
Overview 8
Overview of configuration 16
Overview of error messages 34
Oxy configuration 16
Oxygen-free medium 18
Oxygen sensor, amperometrical 35

P

Package contents 6
Package contents, complete 10
Paraly SW 112 software 31
Parameter setting, data logger 26
Parameter settings (configuration) 16
Ports 14
Power-on 15
Product features 8
Product presentation 8
Protective cover 10

Q

Quickstart guides 7

R

Rating plate 10
RCL, displaying the logger data 28
RCL key 12
Real-time clock 8
Rechargeable battery, Li-ion 13

Recorded data, display 28
Registered trademarks 3
Replacement quiver (accessory) 35
Replacing the batteries 13
Reset to factory settings (Default) 16
Return of products under warranty 3

S

Safety instructions 7
Salinity correction (Salt Correct) 16
Saving the currently measured value 23
Seconds, display 30
Sensoface messages 33
Sensor connection 14
Sensor without temperature detector 21
set key 12
Setting the configuration data 16
Setting the data logger 26
Setting the time and date 30
Setup 16
SHOT (data logger mode) 23
Smiley 33
Software, Paraly SW 112 31
Specifications 36
Specific test report 7
Start address (data logger) 23
Starting the data logger using CONT 27
Starting the data logger using START 27
START, starting the data logger 27
Start-Up 13
Status messages 32
STO, activating the logger 23
STO key 12
Stopping the data logger 29
Structure of data logger 25
Suspending the meter 10
Switching on the meter 15
Symbols in display 15

T

Table of error messages 34
Table view of configuration 16
Technical data 36
Temperature detectors (accessory) 35
Temperature, manual adjustment 21
Temperature probe, connection 14
Trademarks 3
Triangle icons 12

U

USB port for battery charging 13

V

Value-added features 9
Viewing recorded data 28
Viewing the logger data 28

W

Warranty 3

Z

ZERO CAL (calibration) 18

Knick
Elektronische Messgeräte GmbH & Co. KG



Beuckestr. 22
14163 Berlin
Germany

Phone: +49 (0)30 - 801 91 - 0
Fax: +49 (0)30 - 801 91 - 200
Web: www.knick.de
Email: knick@knick.de



085140