

Lovibond® Water Testing Tintometer® Group



XD 7500 (UV-VIS)

Simplify your Work - Save Time and Money!



- Premium optical system with reference beam
- Automatic test recognition with internal barcode reader
- Automatic cuvette type detection
- Support of analytical quality assurance

Part Number: 71307500

Latest technology

The XD 7000 and XD 7500 spectrophotometers are equipped with VIS and UV / VIS reference beam technology.

Automatic test recognition

Each of the more than 150 pre-programmed Lovibond® methods is recognized by barcode as well as the cuvette type.

Analytical quality assurance

Standard procedures for analytical quality assurance support the verification of the photometer, the overall system (including the chemical methodology) and the verification of matrix effects.

Built-in security levels

The devices support the assignment of passwords and the assignment of up to three different levels of rights.

Additional functions

Measurement of transmission and absorption, spectral scan, kinetics analysis, as well as the creation of user-defined methods.

Interfaces for data processing

You want to process your data? With Ethernet, USB B, USB A for external memory, keyboard, barcode scanner and printer different versions are available.

Industry

Chemical Industry | Food and Beverage Industry | Industries Others | Marine Industry | Municipalities | NGO | Oil and Gas | Pharmaceutical Industry | Power and Energy

Application

Boiler Water | Cooling Water | Disinfection Control | Drinking Water Treatment | Galvanization | Pool Water Control | Pool Water Treatment | Raw Water Treatment | Waste Water Treatment

XD 7500 (UV-VIS)

The XD 7000 and XD 7500 are state-of-the-art spectrophotometers. Due to many technical subtleties, such as The automatic test recognition via barcode or over 150 pre-programmed Lovibond® methods makes it ideal for routine or spectral analysis.

Measuring Range

Test Name	Measuring Range	Chemical Method
Alkalinity-m HR T	5 - 500 mg/l CaCO ₃	Acid / Indicator
Alkalinity-m T	5 - 200 mg/l CaCO ₃	Acid / Indicator
Alkalinity-p T	5 - 300 mg/l CaCO ₃	Acid / Indicator
Aluminium PP	0.01 - 0.25 mg/l Al	Eriochrom Cyanine R
Aluminium T	0.01 - 0.3 mg/l Al	Eriochrom Cyanine R
Ammonia HR TT	1.0 - 50 mg/l N	Salicylate
Ammonia LR TT	0.02 - 2.5 mg/l N	Salicylate
Ammonia PP	0.01 - 0.8 mg/l N	Salicylate
Ammonia T	0.02 - 1 mg/l N	Indophenole Blue
Arsenic	0.02 - 0.6 mg/l As	Silver Diethyldithiocarbamate
Boron T	0.1 - 2 mg/l B	Azomethine
Bromine 10 T	0.1 - 3 mg/l Br ₂	DPD
Bromine 50 T	0.05 - 1 mg/l Br ₂	DPD
Bromine PP	0.05 - 4.5 mg/l Br ₂	DPD
Bromine T	0.05 - 13 mg/l Br ₂	DPD
Cadmium M. TT	0.025 - 0.75 mg/l Cd	Cadion
Chloride L (A)	0.5 - 20 mg/l Cl ⁻	Mercury Thiocyanate / Iron Nitrate
Chloride L (B)	5.00 - 60 mg/l Cl ⁻	Iron(III)-thiocyanate
Chloride T	0.5 - 25 mg/l Cl ⁻	Silver Nitrate / Turbidity
Chlorine 10 T	0.1 - 6 mg/l Cl ₂	DPD
Chlorine 50 T	0.02 - 0.5 mg/l Cl ₂ ^{a)}	DPD
Chlorine dioxide 50 T	0.05 - 1 mg/l ClO ₂	DPD / Glycine
Chlorine dioxide PP	0.04 - 3.8 mg/l ClO ₂	DPD
Chlorine dioxide T	0.02 - 11 mg/l ClO ₂	DPD / Glycine
Chlorine HR 10 T	0.1 - 10 mg/l Cl ₂ ^{a)}	DPD
Chlorine HR (KI) T (105)	5 - 200 mg/l Cl ₂	KI / Acid
Chlorine L	0.02 - 4.0 mg/l Cl ₂ ^{a)}	DPD
Chlorine PP	0.02 - 2 mg/l Cl ₂ ^{a)}	DPD
Chlorine PP	0.02 - 3.5 mg/l Cl ₂ ^{a)}	DPD
Chlorine T	0.01 - 6.0 mg/l Cl ₂ ^{a)}	DPD
Chlorite T	0.03 - 2.5 mg/l ClO ₂ ⁻	DPD
Chromium 50 PP	0.005 - 0.5 mg/l Cr ⁶⁺	Diphenylcarbazide
Chromium PP	0.02 - 2 mg/l Cr ⁶⁺	Diphenylcarbazide
COD HR TT	200 - 15000 mg/l COD ^{b)}	Dichromate / H ₂ SO ₄
COD LMR TT	15 - 300 mg/l COD ^{b)}	Dichromate / H ₂ SO ₄
COD LR TT	3 - 150 mg/l COD ^{b)}	Dichromate / H ₂ SO ₄
COD MR TT	20 - 1500 mg/l COD ^{b)}	Dichromate / H ₂ SO ₄
Copper 50 T	0.05 - 1 mg/l Cu ^{a)}	Biquinoline
Copper L	0.05 - 4 mg/l Cu ^{a)}	Bicinchoninate
Copper PP	0.05 - 5 mg/l Cu	Bicinchoninate
Copper T	0.05 - 5 mg/l Cu ^{a)}	Biquinoline
Cyanide 50 L	0.005 - 0.2 mg/l CN ⁻	Pyridine-barbituric Acid
Cyanide L	0.01 - 0.5 mg/l CN ⁻	Pyridine-barbituric Acid
CyA T	10 - 160 mg/l CyA	Melamine
DEHA PP	0.02 - 0.5 mg/l DEHA	PPST
DEHA T (L)	0.02 - 0.5 mg/l DEHA	PPST
Fluoride L	0.05 - 2 mg/l F ⁻	SPADNS
Formaldehyde 10 M. L	1.00 - 5.00 mg/l HCHO	H ₂ SO ₄ / Chromotropic acid
Formaldehyde 50 M. L	0.02 - 1.00 mg/l HCHO	H ₂ SO ₄ / Chromotropic acid
Formaldehyde M. TT	0.1 - 5 mg/l HCHO	H ₂ SO ₄ / Chromotropic acid
H ₂ O ₂ 50 T	0.01 - 0.5 mg/l H ₂ O ₂	DPD / Catalyst
H ₂ O ₂ HR L	40 - 500 mg/l H ₂ O ₂	Titanium Tetrachloride / Acid

Test Name	Measuring Range	Chemical Method
H ₂ O ₂ LR L	1 - 50 mg/l H ₂ O ₂	Titanium Tetrachloride / Acid
H ₂ O ₂ T	0.03 - 3 mg/l H ₂ O ₂	DPD / Catalyst
Hardness Calcium (B) T	20 - 500 mg/l CaCO ₃	Murexide
Hardness Calcium (B) T	50 - 900 mg/l CaCO ₃	Murexide
Hardness total HR T	20 - 500 mg/l CaCO ₃ ⁱ⁾	Metallphthaleine
Hardness total T	2 - 50 mg/l CaCO ₃	Metallphthaleine
Hazen 24	10 - 500 mg/l Pt	(APHA) Platinum Cobalt Standard Method
Hazen 50	10 - 500 mg/l Pt	(APHA) Platinum Cobalt Standard Method
Hydrazine C	0.01 - 0.7 mg/l N ₂ H ₄ ^{o)}	PDMAB
Hydrazine L	0.005 - 0.6 mg/l N ₂ H ₄	Dimethylaminobenzaldehyde
Hydrazine P	0.05 - 0.5 mg/l N ₂ H ₄	Dimethylaminobenzaldehyde
Hypochlorite T	0.2 - 17 % NaOCl	Potassium Iodide
Iron 10 T	0.05 - 1 mg/l Fe	Ferrozine / Thioglycolate
Iron 50 T	0.01 - 0.5 mg/l Fe	Ferrozine / Thioglycolate
Iron (TPTZ) PP	0.02 - 1.8 mg/l Fe	TPTZ
Iron HR L	0.1 - 10 mg/l Fe	Thioglycolate
Iron in Mo PP (224)	0.01 - 1.8 mg/l Fe	TPTZ
Iron LR L (A)	0.03 - 2 mg/l Fe	Ferrozine / Thioglycolate
Iron LR L (B)	0.03 - 2 mg/l Fe	Ferrozine / Thioglycolate
Iron PP	0.01 - 1.5 mg/l Fe ⁹⁾	1,10-Phenanthroline
Iron PP	0.02 - 3 mg/l Fe ⁹⁾	1,10-Phenanthroline
Iron T	0.02 - 1 mg/l Fe	Ferrozine / Thioglycolate
K _{S4.3} T	0.1 - 4 mmol/l K _{S4.3}	Acid / Indicator
Lead 10	0.1 - 5 mg/l Pb	4-(2-Pyridylazo)-resorcin
Lead (A) TT	0.1 - 5 mg/l Pb	4-(2-Pyridylazo)-resorcin
Lead (B) TT	0.1 - 5 mg/l Pb	4-(2-Pyridylazo)-resorcin
Iodine T	0.05 - 3.6 mg/l I	DPD
Manganese HR PP	0.1 - 18 mg/l Mn	Periodate Oxidation
Manganese L	0.05 - 5 mg/l Mn	Formaloxime
Manganese LR PP	0.01 - 0.7 mg/l Mn	PAN
Manganese T	0.2 - 4 mg/l Mn	Formaloxime
Molybdate HR L	1 - 100 mg/l MoO ₄	Thioglycolate
Molybdate HR PP	0.5 - 66 mg/l MoO ₄	Mercaptoacetic Acid
Molybdate LR PP	0.05 - 5 mg/l MoO ₄	Mercaptoacetic Acid
Molybdate T	1 - 50 mg/l MoO ₄	Thioglycolate
Nickel 50 L	0.02 - 1 mg/l Ni	Dimethylglyoxime
Nickel L	0.2 - 7 mg/l Ni	Dimethylglyoxime
Nitrate DMP HR	1.2 - 35 mg/l N	2,6-Dimethylphenole
Nitrate LR TT	0.5 - 14 mg/l N	2,6-Dimethylphenole
Nitrate T	0.08 - 1 mg/l N	Zinc Reduction / NED
Nitrate TT	1 - 30 mg/l N	Chromotropic Acid
Nitrite HR TT	0.3 - 3 mg/l N	Sulfanilic / Naphthylamine
Nitrite LR TT	0.03 - 0.6 mg/l N	Sulfanilic / Naphthylamine
Nitrite PP	0.01 - 0.3 mg/l N	Diazotation
Nitrite T	0.01 - 0.5 mg/l N	N-(1-Naphthyl)-ethylenediamine
Oxygen active T	0.1 - 10 mg/l O ₂	DPD
Oxygen dissolved C	10 - 1100 µg/l O ₂ ^{e)}	Rhodazine D TM
Ozone 50 T	0.02 - 0.5 mg/l O ₃	DPD / Glycine
Ozone PP	0.015 - 2 mg/l O ₃	DPD / Glycine
Ozone T	0.02 - 2 mg/l O ₃	DPD / Glycine
Phenol T	0.1 - 5 mg/l C ₆ H ₅ OH	4-Aminoantipyrine

Test Name	Measuring Range	Chemical Method
PHMB T	2 - 60 mg/l PHMB	Buffer / Indicator
Phosphate h. TT	0.02 - 1.6 mg/l P ^{b)}	Phosphomolybdenum Blue
Phosphate HR C	1.6 - 13 mg/l P ^{c)}	Vanadomolybdate
Phosphate HR L	1.63 - 26.09 mg/l P	Vanadomolybdate
Phosphate HR T	0.33 - 26.09 mg/l P	Vanadomolybdate
Phosphate HR TT	0.98 - 19.57 mg/l P	Vanadomolybdate
Phosphate LR C	0.016 - 1.6 mg/l P ^{c)}	Stannous Chloride
Phosphate LR L	0.033 - 3.261 mg/l P	Phosphomolybic Acid / Ascorbic Acid
Phosphate LR T	0.016 - 1.305 mg/l P	Phosphomolybdenum Blue
Phosphate PP	0.02 - 0.815 mg/l P	Phosphomolybdenum Blue
Phosphate t. TT	0.02 - 1.1 mg/l P ^{b)}	Phosphomolybdenum Blue
Phosphate total HR TT	1.5 - 20 mg/l P ^{b)}	Phosphomolybdenum Blue
Phosphate total LR TT	0.07 - 3 mg/l P ^{b)}	Phosphomolybdenum Blue
Phosphate TT	0.02 - 1.63 mg/l P	Phosphomolybdenum Blue
Phosphonate PP	0.2 - 125 mg/l PO ₄	Persulfate UV Oxidation Method
pH-value HR T	8.0 - 9.6	Thymol Blue
pH value L	6.5 - 8.4	Phenol Red
pH-value LR T	5.2 - 6.8	Bromocresolpurple
pH-value T	6.5 - 8.4	Phenol Red
Polyacrylate L	1 - 30 mg/l Polyacryl	Turbidity
Potassium T	0.7 - 16 mg/l K	Tetraphenylborat Turbidity
SAC 254 nm (344)	0.5 - 50 m ⁻¹	Direct Reading EN ISO 7887:1994
SAC 436 nm	0.5 - 50 m ⁻¹	Direct Reading EN ISO 7887:1994
SAC 525 nm	0.5 - 50 m ⁻¹	Direct Reading EN ISO 7887:1994
SAC 620 nm	0.5 - 50 m ⁻¹	Direct Reading EN ISO 7887:1994
Selenium	0.05 - 2 mg/l Se	3,3'-Diaminobenzidine in Toluene
Silcate T	0.05 - 4 mg/l SiO ₂	Silicomolybdenum Blue
Silicate HR PP	1 - 100 mg/l SiO ₂	Silicomolybdate
Silicate L	0.1 - 8 mg/l SiO ₂	Heteropolyblue
Silicate LR PP	0.05 - 1.6 mg/l SiO ₂	Heteropolyblue
Sulphate PP	5 - 100 mg/l SO ₄ ²⁻	Bariumsulphate Turbidity
Sulphate T	5 - 100 mg/l SO ₄ ²⁻	Bariumsulphate Turbidity
Sulphide T	0.04 - 0.5 mg/l S ²⁻	DPD / Catalyst
Sulphite 10 T	0.1 - 10 mg/l SO ₃	DTNB
Sulphite T	0.1 - 5 mg/l SO ₃	DTNB
Surfactants M. (anion.) TT	0.05 - 2 mg/l SDSA	Methylene Blue
Surfactants M. (cation.) TT	0.05 - 1.5 mg/l CTAB	Disulphine Blue
Surfactants M. (not ionic) TT	0.1 - 7.5 mg/l Triton X-100	TBPE
Suspended solids 24	10 - 750 mg/l TSS	Turbidity / Attenuated Radiation Method
Suspended solids 50	10 - 750 mg/l TSS	Turbidity / Attenuated Radiation Method
TN HR 2 TT	5 - 140 mg/l N ^{b)} ¹⁾	2,6-Dimethylphenole
TN HR TT	5 - 150 mg/l N ^{b)}	Persulphate Digestion
TN LR 2 TT	0.5 - 14 mg/l N ^{b)}	2,6-Dimethylphenole
TN LR TT	0.5 - 25 mg/l N ^{b)}	Persulphate Digestion
TOC HR M. TT	50 - 800 mg/l TOC ^{b)}	H ₂ SO ₄ / Persulphate / Indicator

Test Name	Measuring Range	Chemical Method
TOC LR M. TT	5 - 80 mg/l TOC ^{b)}	H ₂ SO ₄ / Persulphate / Indicator
Triazole PP	1 - 16 mg/l Benzotriazole or Tolyltriazole	Catalyzed UV Digestion
Turbidity 50	5 - 500 FAU	Attenuated Radiation Method
Turbidity 24	10 - 1000 FAU	Attenuated Radiation Method
Urea T	0.1 - 2.5 mg/l Urea	Indophenol / Urease
Zinc L	0.1 - 2.5 mg/l Zn	Zincon / EDTA
Zinc T	0.02 - 1 mg/l Zn	Zincon

Technical Data

Light Source	Xenon flash lamp
Optics	Grid monochromator with reference beam and beam splitter after exit slit
Measurement	Concentration, single and multi-wavelength measurement of absorbance and % transmission, kinetics, spectra
Wavelength Range	190 - 1100 nm (量程)
Wavelength Resolution	1 nm
Wavelength Accuracy	± 1 nm on all Holmium peaks
Wavelength Reproducibility	better than 0,5 nm
Spectral Scope	4 nm
Photometric Range	-3.3 - +3.3 Abs
Photometric Resolution	Absorption: 0.001 ; Transmission: 0.1 %
Photometric Accuracy	0.003 Abs below 0.6 Abs ; 0.5 % from 0.6 to 2.0 Abs
Photometric Reproducibility	0.003 Abs below 0.6 Abs ; 0.5 % from 0.6 to 2.0 Abs
Photometric Linearity	< 1 % up to 2.0 Abs between 340 to 900 nm
Scan Speed	700 - 2000 nm/min.
Stray Light	> 0.05 % Transmission at 340 and 408 nm
Drift	< 0.005 Abs per hour after 15 minutes heat up time
Operation	Membrane Keyboard
Display	7" high contrast colour graphic-display
Suitable Vials	Rectangular Cuvettes 10 mm Rectangular Cuvettes 20 mm Rectangular Cuvettes 50 mm Round Cuvettes 13 mm Round Cuvettes 16 mm Round Cuvettes 24 mm
Automatic Cuvette Recognition	Round cuvettes: 13, 16 and 24 mm ; Rectangular cuvettes: 10, 20 and 50mm
Test Recognition	via internal barcode reader
Interfaces	Ethernet USB B USB A for External Memory Keypad Barcode-Scanner PCL Compatible Printer
Auto-Check	Self-test at each switch-on: test of memory, processor, internal interface, filter lamp and additional calibration of each wavelength
LIMS Compatibility	ASCII, .csv-files
Internal Storage	approx. 5000 data sets (method, user ID, date, result), autostorage function / manual storage function
Security	Password protection possible: 3 different user levels (guest, user, admin)
Power Drain	100 - 240 V, 50/60 Hz
Power Supply	Buffer batteries (4 x AA), power supply unit with cable
Auto - OFF	Yes
Portability	Benchtop

Environmental Conditions	+10 °C to 35 °C (41 °F to 95 °F), ≤ 75 % average humidity within one year 95 %, max. 30 days/year, 85 % all other days
Stock Conditions	-25 °C to +65 °C (-13 °F to 268 °F)
Compliance	CE
Protection Class	IP 30
IP Protection Class	EN 60529
Interference Emission	Class B
Interference Immunity	IEC 61000-4-3
Tolerance Extension	0.008 E
Meter Safety	EC Directive 2014/35/EC EN 61010-1:2010
Languages User Interface	German, English, French, Spanish, Italian, Portuguese, Polish, Indonesian, Russian, Chinese, Japanese, Dutch, Swedish, Norwegian, Czech, Romanian, Macedonian, Slovenian, Hungarian, Turkish, Korean, Vietnamese, Thai, Serbian, Malaysian, Danish, Bulgarian
Languages Quick Start Guide	German, English, French, Spanish, Italian, Portuguese, Polish, Indonesian, Russian, Chinese, Japanese, Dutch, Swedish, Norwegian, Czech, Romanian, Macedonian, Slovenian, Hungarian, Turkish, Korean, Vietnamese, Thai, Serbian, Malaysian, Danish, Bulgarian
Languages Full User Manual	German, English, Spanish, French, Italian, Portuguese, Chinese, Japanese
Dimensions	422 x 195 x 323 mm
Weight	4.5 kg

Title	Part Number
W110/UV/10MM Rectangular cell, Quartz UV	661130
W110/UV/20MM Rectangular cell, Quartz UV	661140
W110/UV/50MM Rectangular cell, Quartz UV	661160
Secondary standard set VIS	711160
12 V-plug connector for XD 7000/7500	71310020
Barcode Hand-held Scanner	71310030
Semimicro cell, 50 mm with lid	71310045

Delivery Scope Text

- 4 batteries (AA)
- 1 power supply cable
- 4 round cuvettes with lid and 1 zero cuvette XD 7x00 (ø 24 mm)
- 1 zero cuvette (ø 16 mm)
- Quickstart guide in 24 languages
- Full user manual in 8 languages

Accessory

Title	Part Number
Batteries (AA), set of 4	1950025
Round cuvette 24 mm, set of 12	197620
Round cuvette 24 mm, set of 5	197629
Cleaning cloth	197635
Round cuvette 16 mm, set of 10	197665
Zero cuvette ø 16 mm for XD 7000/7500	215661
Zero cuvette ø 24 mm for XD 7000/7500	215662
Verification Standard Kit XD 7000/7500	215663
USB cable 3 m	2444482
UV Pen Lamp, 254 nm	400740
Cuvette stand for 6 round cuvettes Ø 24 mm	418951
Cuvette stand for 10 round cuvettes Ø 16 mm	418957
Pipette tips*, 1-5 ml (white) 100 pc.	419066
Pipette tips**, 0,1-1 ml (white), 1000 pc.	419073
Automatic pipette*, 1-5 ml	419076
Automatic pipette**, 0,1-1 ml	419077
W100/OG/10MM Rectangular cell, optical glass	601040
W100/OG/20MM Rectangular cell, optical glass for determination of arsenic	601050
W100/OG/50MM Rectangular cell, optical glass	601070

Tintometer GmbH

Lovibond® Water Testing
Schleefstraße 8-12
44287 Dortmund
Tel.: +49 (0)231/94510-0
Fax: +49 (0)231/94510-30
sales@tintometer.de
www.lovibond.com
Germany

Tintometer China

Room 1001, China Life Tower
16 Chaoyangmenwai Avenue,
Beijing, 100020
Tel.: +86 10 85251111 App. 330
Fax: +86 10 85251001
chinaoffice@tintometer.com
www.lovibond.com
China

Tintometer Inc.

6456 Parkland Drive
Sarasota, FL 34243
Tel: 941.756.6410
Fax: 941.727.9654
sales@tintometer.us
www.lovibond.com
USA

The Tintometer Limited

Lovibond House
Sun Rise Way
Amesbury, SP4 7GR
Tel.: +44 (0)1980 664800
Fax: +44 (0)1980 625412
water.sales@tintometer.com
www.lovibond.com
UK

Tintometer South East Asia

Unit B-3-12, BBT One Boulevard,
Lebu Nilam 2, Bandar Bukit Tinggi,
Klang, 41200, Selangor D.E
Tel.: +60 (0)3 3325 2285/6
Fax: +60 (0)3 3325 2287
lovibond.asia@tintometer.com
www.lovibond.com
Malaysia

Tintometer India Pvt. Ltd.

Door No: 7-2-C-14, 2nd, 3rd & 4th Floor
Sanathnagar Industrial Estate,
Hyderabad, 500018
Telangana
Tel: +91 (0) 40 23883300
Toll Free: 1 800 599 3891/ 3892
indiaoffice@tintometer.com
www.lovibondwater.in
India

Tintometer AG

Hauptstraße 2
5212 Hausen AG
Tel.: +41 (0)56/4422829
Fax: +41 (0)56/4424121
info@tintometer.ch
www.tintometer.ch
Switzerland

Tintometer Brazil

Caixa Postal: 271
CEP: 13201-970
Jundiaí – SP
Tel.: +55 (11) 3230-6410
sales@tintometer.com.br
www.lovibond.com.br
Brazil

Tintometer Spain

Postbox: 24047
08080 Barcelona
Tel.: +34 661 606 770
sales@tintometer.es
www.lovibond.com
Spain